

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N: 130135

Supervisor/Operator Performing the Verification Procedure:

Name: Trp. J. Santora

ID# 2276

Date: 04/16/2011

A

Agency: Alaska State Troopers

Phone # 656-1233

Instrument Location: Galena Police Department

B

Alco S/N: 78855

Target Value: .081

High Pressure: 950

Alco Test Values

.082  
1<sup>st</sup> Alco

.082  
2<sup>nd</sup> Alco

Signature

[Signature]  
(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2010.

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

APRIL 16, 2011

CRF 4/21/11

OPERATOR'S NAME:  
SANTORA JEANNINE L  
OPERATOR'S NUMBER: 2276  
SUBJECT'S LAST NAME:  
VERIFICATION  
SUBJECT'S FIRST NAME/NI :  
OF CAL  
O.L. #: 0000000  
DEPT/AGENCY: GALO  
CASE/REPORT:  
TEST TYPE: U  
ALCO TARGET VALUE: .001  
ALCO S/N: 78855

### --- BREATH ANALYSIS ---

.001 ADJUSTED FOR 30.43 in  
ALCO TARGET .002 15:19  
BLANK TEST .000 15:20  
INTERNAL STANDARD VERIFIED 15:20  
ALCO TO 30.43 in .002 15:21  
BLANK TEST .000 15:22  
SUBJECT SAMPLE .000 15:23  
BLANK TEST .000 15:24  
ALCO TO 30.43 in .002 15:24  
BLANK TEST .000 15:25

CRF 4/21/11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130135

APRIL 16, 2011  
TIME 15:29

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 43c  
BAROMETER: 30.43 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

"0022"(:++-.0127456789:K=>00200000  
HIJKLMNOPQRSTUVWXYZ~!@. 'abcde fghijklmnop  
qrstuvwxyz{ }~!@